

**Rajdhani College**  
**(University of Delhi)**  
Mahatma Gandhi Marg, Raja garden,  
New Delhi -110015

RC/RG/113/1699

Dated: 31.08.2017


**Corrigendum**

With reference to the notice No. RC/RG/113/1413 dated 12/08/17 given on college website and university website for Benchtop Nitrogen Box/purge glove box it is informed to all concern that the specification for the equipment are modified as

S.No	Specification	Description
	Glove box	<b>Internal dimensions (inches):</b> 34 (W) x 30 (H) x 24 (D) ( $\pm$ 15%) • <b>Material of construction:</b> Stainless steel • <b>Environment:</b> Nitrogen, Argon & Helium


The last date for sending quotations is extended to 08/09/17. The quotations with specifications other than notified will also be considered. Other terms and conditions are same as mentioned in notice No. RC/RG/113/1413.

  
(Dr. Pawan Kumar)  
Principal Investigator

  
(Dr. Rajesh Giri)  
ACTING Principal

Copy to:-

1. Bursar/S.O.(Accounts)
2. AO
3. Principal Investigator
- ✓ 4. Website

  
(Dr. Pawan Kumar)  
Principal Investigator

  
(Dr. Rajesh Giri)  
ACTING Principal